


Assay Class: High Sensitivity DNA Assay
Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
Modified: 9/19/2016 11:01:14 AM

Electrophoresis File Run Summary



Instrument Information:

Instrument Name: DE13701086 Firmware: C.01.069
Serial#: DE13701086 Type: G2938B

Assay Information:



Assay Origin Path: C:\Program Files\Agilent\2100 bioanalyzer\2100 expert\assays\dsDNA\High Sensitivity DNA.xsy
Assay Class: High Sensitivity DNA Assay
Version: 1.03
Assay Comments: Copyright © 2003-2010 Agilent Technologies

Chip Information:

Chip Lot #:
Reagent Kit Lot #:
Chip Comments:

KA-1 (1:20)

KA-2 (1:20)



Assay Class: High Sensitivity DNA Assay
Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
Modified: 9/19/2016 11:01:14 AM

Electrophoresis File Run Summary (Chip Summary)

Sample Name	Sample Comment	Rest. Digest	Status	Observation	Result Label	Result Color
KA-1 (1:20)		<input type="checkbox"/>	✓			
KA-2 (1:20)		<input type="checkbox"/>	✓			
Ladder		<input type="checkbox"/>	✓			

Chip Lot # **Reagent Kit Lot #**

Chip Comments :

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
Modified: 9/19/2016 11:01:14 AM

Electrophoresis Assay Details

General Analysis Settings

Number of Available Sample and Ladder Wells (Max.) : 12
Minimum Visible Range [s] : 32
Maximum Visible Range [s] : 138
Start Analysis Time Range [s] : 33
End Analysis Time Range [s] : 137.5
Ladder Concentration [pg/μl] : 1950
Uses Standard Area for Ladder Fragments
Lower Marker Concentration [pg/μl] : 125
Upper Marker Concentration [pg/μl] : 75
Used Upper Marker for Quantitation
Standard Curve Fit is Point to Point
Show Data Aligned to Lower and Upper Marker

Integrator Settings

Integration Start Time [s] : 33.05
Integration End Time [s] : 137
Slope Threshold : 0.8
Height Threshold [FU] : 5
Area Threshold : 0.1
Width Threshold [s] : 0.6
Baseline Plateau [s] : 0.5

Filter Settings

Filter Width [s] : 0.5
Polynomial Order : 4


Ladder

Ladder Peak	Size	Area
1	35	160
2	50	210
3	100	208
4	150	221
5	200	242
6	300	270
7	400	305
8	500	306
9	600	336
10	700	321
11	1000	366
12	2000	413
13	3000	411
14	7000	400
15	10380	214

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Electropherogram Summary



Overall Results for Ladder

Noise: 0.2


Peak table for Ladder

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations	Aligned Migration Time [s]
1	35	125.00	5,411.3	Lower Marker	43.00
2	50	150.00	4,545.5	Ladder Peak	45.46
3	100	150.00	2,272.7	Ladder Peak	51.26
4	150	150.00	1,515.2	Ladder Peak	56.13
5	200	150.00	1,136.4	Ladder Peak	60.96
6	300	150.00	757.6	Ladder Peak	70.30
7	400	150.00	568.2	Ladder Peak	78.22
8	500	150.00	454.5	Ladder Peak	83.93
9	600	150.00	378.8	Ladder Peak	88.65
10	700	150.00	324.7	Ladder Peak	91.65
11	1,000	150.00	227.3	Ladder Peak	95.78
12	2,000	150.00	113.6	Ladder Peak	101.69
13	3,000	150.00	75.8	Ladder Peak	104.64
14	7,000	150.00	32.5	Ladder Peak	109.61
15	10,380	75.00	10.9	Upper Marker	113.00

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Electropherogram Summary Continued ...



Overall Results for sample 9 : KA-1 (1:20)

Number of peaks found: 31 Corr. Area 1: 2,694.8
 Noise: 0.6

Peak table for sample 9 : KA-1 (1:20)

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations	Aligned Migration Time [s]
1	35	125.00	5,411.3	Lower Marker	43.00
2	49	20.08	625.9		45.23
3	106	84.99	1,212.4		51.87
4	129	103.92	1,219.5		54.10
5	141	71.26	765.2		55.27
6	152	47.46	473.0		56.33
7	165	80.01	733.6		57.61
8	185	125.29	1,025.6		59.52
9	194	47.96	374.8		60.37
10	219	70.04	484.0		62.76
11	244	161.68	1,005.1		65.04
12	269	140.03	789.5		67.38
13	292	134.81	699.5		69.56
14	303	72.35	362.2		70.51
15	321	114.05	538.7		71.95
16	342	138.40	614.0		73.59
17	380	104.05	415.1		76.62
18	395	57.43	220.6		77.79
19	409	69.82	258.5		78.74
20	441	57.02	196.0		80.55
21	499	94.74	287.4		83.90
22	551	65.39	179.8		86.34
23	589	58.85	151.3		88.14
24	654	65.08	150.8		90.27
25	700	64.53	139.7		91.65

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Electropherogram Summary Continued ...

... Peak table for sample 9 : KA-1 (1:20)

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations	Aligned Migration Time [s]
26	835	47.26	85.8		93.51
27	870	106.03	184.8		93.99
28	1,029	87.14	128.4		95.95
29	1,397	40.86	44.3		98.13
30	1,739	62.79	54.7		100.15
31	4,773	3.44	1.1		106.84
32	7,207	2.65	0.6		109.81
33	10,380	75.00	10.9	Upper Marker	113.00


Region table for sample 9 : KA-1 (1:20)

From [s]	To [s]	Average Size [bp]	Corr. Area	Molarity [pmol/l]	Co Conc. lor [pg/μl]	% of Total	Size distribution in CV [%]
48.43	108.54	591	2,694.8	17,347.4	3,431.50	97	95.8

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Electropherogram Summary Continued ...



Overall Results for sample 10 : KA-2 (1:20)

Number of peaks found: 28 Corr. Area 1: 3,049.0
 Noise: 0.3

Peak table for sample 10 : KA-2 (1:20)

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations	Aligned Migration Time [s]
1	35	125.00	5,411.3	Lower Marker	43.00
2	160	355.88	3,365.9		57.12
3	217	399.31	2,791.6		62.52
4	233	146.45	953.3		64.02
5	248	131.75	804.3		65.46
6	269	183.67	1,035.3		67.39
7	282	134.81	724.4		68.61
8	294	96.72	498.5		69.74
9	311	97.74	476.0		71.18
10	322	97.81	460.4		72.04
11	332	86.57	395.0		72.84
12	341	80.47	357.8		73.53
13	356	90.67	386.2		74.71
14	406	88.43	330.0		78.56
15	429	110.24	389.0		79.90
16	452	72.26	242.3		81.18
17	464	82.65	269.8		81.88
18	509	96.44	287.3		84.34
19	552	91.11	250.2		86.37
20	572	122.64	324.8		87.33
21	599	69.71	176.3		88.61
22	699	95.25	206.6		91.61
23	883	75.90	130.2		94.18
24	973	103.09	160.6		95.41
25	1,235	148.63	182.3		97.17

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Electropherogram Summary Continued ...

... Peak table for sample 10 : KA-2 (1:20)

Peak	Size [bp]	Conc. [pg/ μ l]	Molarity [pmol/l]	Observations	Aligned Migration Time [s]
26	1,815	103.11	86.1		100.59
27	4,566	9.47	3.1		106.58
28	6,934	4.10	0.9		109.52
29	7,771	4.02	0.8		110.38
30	10,380	75.00	10.9	Upper Marker	113.00


Region table for sample 10 : KA-2 (1:20)

From [s]	To [s]	Average Size [bp]	Corr. Area	Molarity [pmol/l]	Co Conc. lor [pg/ μ l]	% of Total	Size distribution in CV [%]
46.53	109.52	645	3,049.0	22,132.9	4,503.70	99	100.0

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
Modified: 9/19/2016 11:01:14 AM

Gel Image



Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...yzer\2100 expert\data\2016-09-19\2016-09-19_001_KA1_KA2.xad

Created: 9/19/2016 10:14:11 AM
 Modified: 9/19/2016 11:01:14 AM

Run Logbook

Description	Number	Source	Category	Sub Category	Time	Time Zone	User	Host
Run ended on port 1 (Number of wells acquired: 12)		Instrument	Run		9/19/2016 10:55:28 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Run started on port 1 (File: C:\Program Files\Agilent\2100 bioanalyzer\2100 expert\data\2016-09-19\2016-09-19_001.xad)		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Product Number : G2938B		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Name :		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Vendor : Agilent Technologies		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Serial# : DE13701086		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Firmware : C.01.069		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Cartridge : Electrode		Instrument	Run		9/19/2016 10:14:16 AM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1