


Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
Modified: 3/24/2014 4:04:28 PM

Electrophoresis File Run Summary



Instrument Information:


Instrument Name: DE13701086 Firmware: C.01.069
Serial#: DE13701086 Type: G2938B

Assay Information:

Assay Origin Path: C:\Program Files\Agilent\2100 bioanalyzer\2100 expert\assays\dsDNA\High Sensitivity DNA.xsy
Assay Class: High Sensitivity DNA Assay
Version: 1.03
Assay Comments: Copyright © 2003-2010 Agilent Technologies

Chip Information:

Chip Lot #:
Reagent Kit Lot #:
Chip Comments:



Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electrophoresis File Run Summary (Chip Summary)

Sample Name	Sample Comment	Rest. Digest	Status	Observation	Result Label	Result Color
ELVS1		<input type="checkbox"/>	✓			
ELVS2		<input type="checkbox"/>	✓			
ELVS3		<input type="checkbox"/>	✓			
ELVS4		<input type="checkbox"/>	✓			
ELVS5		<input type="checkbox"/>	✓			
ELVS6		<input type="checkbox"/>	✓			
ELVS7		<input type="checkbox"/>	✓			
ELVS8		<input type="checkbox"/>	✓			
sample 9		<input type="checkbox"/>				
sample 10		<input type="checkbox"/>				
sample 11		<input type="checkbox"/>				
Ladder		<input type="checkbox"/>	✓			

Chip Lot #

Reagent Kit Lot #

Chip Comments :

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
Modified: 3/24/2014 4:04:28 PM

Electrophoresis Assay Details

General Analysis Settings

Number of Available Sample and Ladder Wells (Max.) : 12
Minimum Visible Range [s] : 32
Maximum Visible Range [s] : 138
Start Analysis Time Range [s] : 33
End Analysis Time Range [s] : 137.5
Ladder Concentration [pg/μl] : 1950
Uses Standard Area for Ladder Fragments
Lower Marker Concentration [pg/μl] : 125
Upper Marker Concentration [pg/μl] : 75
Used Upper Marker for Quantitation
Standard Curve Fit is Point to Point
Show Data Aligned to Lower and Upper Marker

Integrator Settings

Integration Start Time [s] : 33.05
Integration End Time [s] : 137
Slope Threshold : 0.8
Height Threshold [FU] : 5
Area Threshold : 0.1
Width Threshold [s] : 0.6
Baseline Plateau [s] : 0.5

Filter Settings

Filter Width [s] : 0.5
Polynomial Order : 4


Ladder

Ladder Peak	Size	Area
1	35	160
2	50	210
3	100	208
4	150	221
5	200	242
6	300	270
7	400	305
8	500	306
9	600	336
10	700	321
11	1000	366
12	2000	413
13	3000	411
14	7000	400
15	10380	214

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary



Overall Results for Ladder

Noise: 0.2


Peak table for Ladder

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	45.28	50	150.00	4,545.5	Ladder Peak
3	50.58	100	150.00	2,272.7	Ladder Peak
4	55.14	150	150.00	1,515.2	Ladder Peak
5	59.62	200	150.00	1,136.4	Ladder Peak
6	68.51	300	150.00	757.6	Ladder Peak
7	76.66	400	150.00	568.2	Ladder Peak
8	82.36	500	150.00	454.5	Ladder Peak
9	87.00	600	150.00	378.8	Ladder Peak
10	90.26	700	150.00	324.7	Ladder Peak
11	94.34	1,000	150.00	227.3	Ladder Peak
12	101.06	2,000	150.00	113.6	Ladder Peak
13	104.44	3,000	150.00	75.8	Ladder Peak
14	109.50	7,000	150.00	32.5	Ladder Peak
15	113.00	10,380	75.00	10.9	Upper Marker

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 1 : ELVS1

Height Threshold [FU] : 45

Overall Results for sample 1 : ELVS1

Number of peaks found: 5 Corr. Area 1: 4,619.3
 Noise: 0.2

Peak table for sample 1 : ELVS1

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	67.19	285	2,976.49	15,815.9	
3	67.76	292	554.94	2,883.9	
4	68.57	301	377.88	1,903.6	
5	69.63	314	3,249.65	15,693.6	
6	76.13	394	511.77	1,970.3	
7	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 1 : ELVS1

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
189	487	317	38,813.4	7,830.34	4,619.3	100	16.4

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 2 : ELVS2

Height Threshold [FU] : 45

Overall Results for sample 2 : ELVS2

Number of peaks found: 6 Corr. Area 1: 3,528.1
 Noise: 0.5

Peak table for sample 2 : ELVS2

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	65.39	265	2,393.09	13,685.9	
3	66.48	277	3,511.15	19,191.6	
4	72.41	348	431.68	1,879.9	
5	73.18	357	401.64	1,703.0	
6	74.47	373	536.21	2,177.1	
7	76.65	400	177.97	674.3	
8	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 2 : ELVS2

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
185	460	301	35,696.2	6,873.29	3,528.1	100	16.1

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 3 : ELVS3

Height Threshold [FU] : 30

Overall Results for sample 3 : ELVS3

Number of peaks found: 4 Corr. Area 1: 6,048.9
 Noise: 0.5

Peak table for sample 3 : ELVS3

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	65.79	269	3,657.97	20,575.1	
3	66.73	280	1,018.18	5,509.5	
4	67.06	284	6,930.14	37,011.5	
5	75.15	381	838.92	3,331.9	
6	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 3 : ELVS3

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
190	472	309	64,312.3	12,681.64	6,048.9	100	16.5

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Overall Results for sample 4 : ELVS4

Number of peaks found: 3 Corr. Area 1: 3,651.5
 Noise: 0.8

Peak table for sample 4 : ELVS4

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	42.00	31	0.00	0.0	
2	43.00	35	125.00	5,411.3	Lower Marker
3	44.21	43	24.10	849.6	
4	66.79	281	6,891.87	37,198.8	
5	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 4 : ELVS4

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
182	475	303	34,033.7	6,594.26	3,651.5	99	16.2

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 5 : ELVS5

Height Threshold [FU] : 45

Overall Results for sample 5 : ELVS5

Number of peaks found: 4 Corr. Area 1: 5,968.7
 Noise: 0.5

Peak table for sample 5 : ELVS5

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	67.60	290	2,377.68	12,433.4	
3	68.79	303	397.56	1,985.1	
4	69.26	309	1,065.56	5,221.6	
5	70.74	327	4,024.71	18,621.8	
6	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 5 : ELVS5

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
196	540	332	37,520.4	7,928.92	5,968.7	99	16.8

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 6 : ELVS6

Height Threshold [FU] : 40

Overall Results for sample 6 : ELVS6

Number of peaks found: 2 Corr. Area 1: 3,604.2
 Noise: 0.5

Peak table for sample 6 : ELVS6

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	65.86	270	2,593.86	14,544.9	
3	66.89	282	6,565.65	35,296.9	
4	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 6 : ELVS6

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. for Area	% of Total	Size distribution in CV [%]
182	484	307	43,531.9	8,522.38	3,604.2	99	16.5

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 7 : ELVS7

Height Threshold [FU] : 45

Overall Results for sample 7 : ELVS7

Number of peaks found: 2 Corr. Area 1: 4,731.7
 Noise: 0.4

Peak table for sample 7 : ELVS7

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	67.58	290	2,581.18	13,507.5	
3	68.67	302	5,166.55	25,916.0	
4	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 7 : ELVS7

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
193	536	328	36,704.3	7,645.02	4,731.7	99	17.0

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Electropherogram Summary Continued ...



Setpoint Deviations for sample 8 : ELVS8

Height Threshold [FU] : 40

Overall Results for sample 8 : ELVS8

Number of peaks found: 1 Corr. Area 1: 4,453.4
 Noise: 0.3

Peak table for sample 8 : ELVS8

Peak	Aligned Migration Time [s]	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	43.00	35	125.00	5,411.3	Lower Marker
2	66.90	282	8,211.39	44,139.2	
3	113.00	10,380	75.00	10.9	Upper Marker


Region table for sample 8 : ELVS8

From [bp]	To [bp]	Average Size [bp]	Molarity [pmol/l]	Conc. [pg/μl]	Co Corr. lor Area	% of Total	Size distribution in CV [%]
188	496	312	38,291.6	7,600.13	4,453.4	99	16.8

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
Modified: 3/24/2014 4:04:28 PM

Gel Image




Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
Modified: 3/24/2014 4:04:28 PM

Curves

Standard Curve



Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
Modified: 3/24/2014 4:04:28 PM

Invalid Samples

Sample 9 has not been run, no results available.

Sample 10 has not been run, no results available.

Sample 11 has not been run, no results available.

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad

Created: 3/24/2014 3:28:08 PM
 Modified: 3/24/2014 4:04:28 PM

Run Logbook

Description	Number	Source	Category	Sub Category	Time	Time Zone	User	Host
Run ended on port 1 (Number of wells acquired: 9)		Instrument	Run		3/24/2014 4:00:52 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Run started on port 1 (File: C:\Documents and Settings\Bioanalyzer\2014-03-24\2014-03-24_006.xad)		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Product Number : G2938B		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Name :		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Vendor : Agilent Technologies		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Serial# : DE13701086		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Firmware : C.01.069		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1
Cartridge : Electrode		Instrument	Run		3/24/2014 3:28:13 PM	(GMT --07:00) Pacific Standard Time	UC Davis	D8XSMGH1