


Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
Modified: 11/24/2014 4:39:03 PM

Electrophoresis File Run Summary



Instrument Information:










Instrument Name: DE13701086 Firmware: C.01.069
Serial#: DE13701086 Type: G2938B

Assay Information:

Assay Origin Path: C:\Program Files\Agilent\2100 bioanalyzer\2100 expert\assays\dsDNA\High Sensitivity DNA.xsy
Assay Class: High Sensitivity DNA Assay
Version: 1.03
Assay Comments: Copyright © 2003-2010 Agilent Technologies

Chip Information:

Chip Lot #: Reagent Kit Lot #: Chip Comments:



Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electrophoresis File Run Summary (Chip Summary)

Sample Name	Sample Comment	Rest. Digest	Status	Observation	Result Label	Result Color
sample E2	untreated-lib prep	<input type="checkbox"/>	✓			
sample C7	untreated-lib prep	<input type="checkbox"/>	✓			
sample F2	untreated-lib prep	<input type="checkbox"/>	⚠			
sample F3	untreated-lib prep	<input type="checkbox"/>	✓			
sample C1	treated-lib prep	<input type="checkbox"/>	✓			
sample D1	treated-lib prep	<input type="checkbox"/>	✓			
Sample E1	treated-lib prep	<input type="checkbox"/>	⚠			
sample F1	treated-lib prep	<input type="checkbox"/>	✓			
sample D1	Frag. non-lib prep	<input type="checkbox"/>	✓			
sample 10		<input type="checkbox"/>				
sample 11		<input type="checkbox"/>				
Ladder		<input type="checkbox"/>	✓			

Chip Lot #

Reagent Kit Lot #

Chip Comments :

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
Modified: 11/24/2014 4:39:03 PM

Electrophoresis Assay Details

General Analysis Settings

Number of Available Sample and Ladder Wells (Max.) : 12
Minimum Visible Range [s] : 32
Maximum Visible Range [s] : 138
Start Analysis Time Range [s] : 33
End Analysis Time Range [s] : 137.5
Ladder Concentration [pg/μl] : 1950
Uses Standard Area for Ladder Fragments
Lower Marker Concentration [pg/μl] : 125
Upper Marker Concentration [pg/μl] : 75
Used Upper Marker for Quantitation
Standard Curve Fit is Point to Point
Show Data Aligned to Lower and Upper Marker

Integrator Settings

Integration Start Time [s] : 33.05
Integration End Time [s] : 137
Slope Threshold : 0.8
Height Threshold [FU] : 5
Area Threshold : 0.1
Width Threshold [s] : 0.6
Baseline Plateau [s] : 0.5

Filter Settings

Filter Width [s] : 0.5
Polynomial Order : 4


Ladder

Ladder Peak	Size	Area
1	35	160
2	50	210
3	100	208
4	150	221
5	200	242
6	300	270
7	400	305
8	500	306
9	600	336
10	700	321
11	1000	366
12	2000	413
13	3000	411
14	7000	400
15	10380	214

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary



Overall Results for Ladder

Noise: 0.1


Peak table for Ladder

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	50	150.00	4,545.5	Ladder Peak
3	100	150.00	2,272.7	Ladder Peak
4	150	150.00	1,515.2	Ladder Peak
5	200	150.00	1,136.4	Ladder Peak
6	300	150.00	757.6	Ladder Peak
7	400	150.00	568.2	Ladder Peak
8	500	150.00	454.5	Ladder Peak
9	600	150.00	378.8	Ladder Peak
10	700	150.00	324.7	Ladder Peak
11	1,000	150.00	227.3	Ladder Peak
12	2,000	150.00	113.6	Ladder Peak
13	3,000	150.00	75.8	Ladder Peak
14	7,000	150.00	32.5	Ladder Peak
15	10,380	75.00	10.9	Upper Marker

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 1 : sample E2

Number of peaks found: 1 Corr. Area 1: 24.7
 Noise: 0.2

Peak table for sample 1 : sample E2

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	328	2.87	13.2	
3	10,380	75.00	10.9	Upper Marker


Region table for sample 1 : sample E2

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	258	1,000	24.7	72	12.5	30.55	183.0	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 2 : sample C7

Number of peaks found: 1 Corr. Area 1: 0.0
 Noise: 0.1

Peak table for sample 2 : sample C7

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	45	35.88	1,195.9	
3	10,380	75.00	10.9	Upper Marker


Region table for sample 2 : sample C7

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	0	1,000	0.0	0	0.0	0.00	0.0	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 3 : sample F2

Number of peaks found: 0 Corr. Area 1: 151.4
 Noise: 0.1

Peak table for sample 3 : sample F2

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	10,380	75.00	10.9	Upper Marker


Region table for sample 3 : sample F2

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	606	1,000	151.4	59	15.6	156.14	441.6	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 4 : sample F3

Number of peaks found: 1 Corr. Area 1: 53.9
 Noise: 0.2

Peak table for sample 4 : sample F3

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	46	17.12	566.4	
3	10,380	75.00	10.9	Upper Marker


Region table for sample 4 : sample F3

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	446	1,000	53.9	41	39.4	53.56	230.7	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 5 : sample C1

Number of peaks found: 1 Corr. Area 1: 14.7
 Noise: 0.2

Peak table for sample 5 : sample C1

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	45	30.86	1,033.5	
3	10,380	75.00	10.9	Upper Marker


Region table for sample 5 : sample C1

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	387	1,000	14.7	17	49.7	15.23	80.6	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 6 : sample D1

Number of peaks found: 2 Corr. Area 1: 40.1
 Noise: 0.1

Peak table for sample 6 : sample D1

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	45	26.07	874.2	
3	10,380	75.00	10.9	Upper Marker
4	12,464	0.00	0.0	


Region table for sample 6 : sample D1

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	456	1,000	40.1	36	30.5	60.39	233.5	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 7 : Sample E1

Number of peaks found: 1 Corr. Area 1: 49.3
 Noise: 0.1

Peak table for sample 7 : Sample E1

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	45	38.85	1,303.2	
3	10,380	75.00	10.9	Upper Marker


Region table for sample 7 : Sample E1

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
200	532	1,000	49.3	9	31.9	51.54	174.5	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 8 : sample F1

Number of peaks found: 2 Corr. Area 1: 49.8
 Noise: 0.1

Peak table for sample 8 : sample F1

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	45	45.66	1,525.5	
3	10,380	75.00	10.9	Upper Marker
4	12,809	0.00	0.0	


Region table for sample 8 : sample F1

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
297	502	1,000	49.8	50	23.7	58.15	188.2	Blue

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Electropherogram Summary Continued ...



Overall Results for sample 9 : sample D1

Number of peaks found: 5 Corr. Area 1: 1,446.9
 Noise: 0.1

Peak table for sample 9 : sample D1

Peak	Size [bp]	Conc. [pg/μl]	Molarity [pmol/l]	Observations
1	35	125.00	5,411.3	Lower Marker
2	253	13.00	77.8	
3	290	12.82	66.9	
4	324	23.87	111.8	
5	358	10.40	44.0	
6	403	9.05	34.1	
7	10,380	75.00	10.9	Upper Marker


Region table for sample 9 : sample D1

From [bp]	Average Size [bp]	To [bp]	Corr. Area	% of Total	Size distribution in CV [%]	Conc. [pg/μl]	Molarity [pmol/l]	Color
49	346	2,564	1,446.9	97	80.9	2,682.44	20,357.4	Blue

Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
Modified: 11/24/2014 4:39:03 PM

Gel Image



Assay Class: High Sensitivity DNA Assay
Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
Modified: 11/24/2014 4:39:03 PM

Invalid Samples

Sample 10 has not been run, no results available.

Sample 11 has not been run, no results available.

Assay Class: High Sensitivity DNA Assay
 Data Path: C:\...ents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad

Created: 11/24/2014 3:56:22 PM
 Modified: 11/24/2014 4:39:03 PM

Run Logbook

Description	Number	Source	Category	Sub Category	Time	Time Zone	User	Host
Run ended on port 1 (Number of wells acquired: 10)		Instrument	Run		11/24/2014 4:31:58 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Instrument error occurred on port 1, Optical signal too high (1605h)	559	Instrument	Run	Sample 7	11/24/2014 4:24:39 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Instrument error occurred on port 1, Optical signal too high (1605h)	559	Instrument	Run	Sample 3	11/24/2014 4:14:23 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Run started on port 1 (File: C:\Documents and Settings\Bioanalyzer\2014-11-24\2014-11-24_003.xad)		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Product Number : G2938B		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Name :		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Vendor : Agilent Technologies		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Serial# : DE13701086		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Firmware : C.01.069		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1
Cartridge : Electrode		Instrument	Run		11/24/2014 3:56:28 PM	(GMT --08:00) Pacific Standard Time	UC Davis	D8XSMGH1